

## Test Report

No.: CANHL24000185701

Date: Jan 10, 2024

Page 1 of 3

Client Name: Maymom LLC

Client Address: 20-7 Aly 19 LN 333, Zhongcheng Rd Zhunan Township, Miaoli County, Taiwan

Sample Description: Breast Pump Flange

Buyer: Amazon, US, UK, China, Canada, EU

Manufacturer: Maymom LLC

Client Ref. Information: Material: PP+TPR; Model: MP012F-F19 Batch: R20231121-10

The above sample(s) and information were provided by the client.

SGS No.: XMHL2401000058CW

Sample Receiving Date: Jan 03, 2024

Testing Period: Jan 03, 2024 ~ Jan 10, 2024

Test Requested: Select test(s) as requested by the client.

Testing Performed: Follow selected test(s) as requested by client.

	Test Requirement	Conclusion
1	Regulation (EC) No 1935/2004 of the European Parliament and of the Council of 27 October 2004, Council of Europe Resolution AP (2004) 4 - Overall Migration	Pass

### Kim Cai

Kim Cai

Approved Signatory

Scan to see the report



CANHL24000185701  
Verification:  
check.sgsonline.com.cn



SGS-CS Standards Technical Services Co., Ltd.  
Guangzhou Branch Technical Services Co., Ltd. Chemical Laboratory

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.  
**Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com**

No.198, Kezhu Road, Science City, Economic & Technological Development Area, Guangzhou, Guangdong, China 510663  
中国·广东·广州高新技术产业开发区科学城科珠路198号 邮编: 510663

t (86-20) 82155555 www.sgsgroup.com.cn  
t (86-20) 82155555 sgs.china@sgs.com

**Test Result(s):**

**Test Part Description:**

SN ID	Sample No.	SGS Sample ID	Description	Material (claimed by the client)
SN1	A1	CAN24-0001857-0001.C001	White soft part	TPR

**Remarks:**

- (1) mg/dm<sup>2</sup> = milligram per square decimeter  
 mg/inch<sup>2</sup> = milligram per square inches  
 µg/cm<sup>2</sup> = microgram per square centimeter  
 mg/L = milligram per litre  
 mg/kg = milligram per kilogram
- (2) °C = degree Celsius  
 °F = degree Fahrenheit
- (3) <= less than
- (4) RL = Reporting Limit
- (5) ND = Not Detected (< RL).

**The ratio of surface area to volume**

SGS Sample ID	Overall Migration
CAN24-0001857-0001.C001	6.0 dm <sup>2</sup> /kg

**Regulation (EC) No 1935/2004 of the European Parliament and of the Council of 27 October 2004, Council of Europe Resolution AP (2004) 4 - Overall Migration**

**Test Method:** With reference to Commission Regulation (EU) No 10/2011 of 14 January 2011 Annex III and Annex V for selection of condition and EN 1186-1:2002 for selection of test methods; EN 1186-3:2022 Test methods for overall migration in evaporable simulants

Test Condition	
50% Ethanol(V/V)	
A1	70°C,2 hrs

Simulant used	Limit	Unit(s)	RL	A1 (3 <sup>rd</sup> Migration)	Comment
50% Ethanol(V/V)	60	mg/kg	20	ND	Pass

**Notes:**

- (1) Analytical tolerance of evaporable simulants is 2 mg/dm<sup>2</sup> or 12 mg/kg, analytical tolerance of olive oil simulant is 3 mg/dm<sup>2</sup> or 20 mg/kg.
  - (2) Test condition & simulant were specified by client.
- Unless otherwise stated, the decision rule for conformity reporting is based on Binary Statement for Simple Acceptance Rule (w=0) stated in ILAC-G8:09/2019.



## Test Report

No.: CANHL24000185701

Date: Jan 10, 2024

Page 3 of 3

### Sample Photo:



SGS authenticate the photo on original report only

\*\*\* End of Report \*\*\*



SGS-CSTC Standards Technical Services Co., Ltd.  
Guangzhou Branch Technical Services Co., Ltd. Guangzhou Branch

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

No.198, Kezhu Road, Science City, Economic & Technological Development Area, Guangzhou, Guangdong, China 510663  
中国·广东·广州高新技术产业开发区科学城科珠路198号 邮编: 510663

t (86-20) 82155555 [www.sgsgroup.com.cn](http://www.sgsgroup.com.cn)  
t (86-20) 82155555 [sgs.china@sgs.com](mailto:sgs.china@sgs.com)